

Certificate of Analysis

Product Information:

Orion Number	900004
Description	REF ELEC FILL SOLN, 5 X 60 ML
Lot Number	HR1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	2.00 M +/- 0.10 M KCL

Actual Analysis:


<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	2.02	Moles/Liter Potassium Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 09/12/03

Approved By:



Quality Control Supervisor