

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	9107CB
<b>Description</b>	PH 7 BUFFER,CUBITAINER 19L
<b>Lot Number</b>	HQ3
<b>Expiration Date</b>	10/05
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	7.00 +/- 0.01 pH units

## Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	7.012	pH

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b

**Certificate Date:** 10/22/03

**Approved By:**



**Quality Control Supervisor**