

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	910110
<b>Description</b>	# 10.01 BUFFER, 475 ML
<b>Lot Number</b>	HR1
<b>Expiration Date</b>	9/05
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	10.01 +/- 0.02

## Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	10.020	pH

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Carbonate	192	b
Sodium Bicarbonate	191	b
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f

**Certificate Date:** 09/17/03

**Approved By:**



**Quality Control Supervisor**