

Certificate of Analysis

Product Information:

Orion Number	654201
Description	0.05M HYAMINE TITRANT, 475ML
Lot Number	HP1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	0.050 M +/- 0.0025 M Hyamine 1622R

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	0.0508	Moles/Liter Hyamine 1622

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 11/07/03

Approved By:



Quality Control Supervisor