

Certificate of Analysis

Product Information:

Orion Number	841109
Description	SODIUM KA STD W/ISA, 1000 PPM
Lot Number	BY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Flame Photometer
Acceptance Range	1000 ppm +/- 5 ppm Sodium as Sodium Chloride

Actual Analysis:


<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	998.5	ppm Sodium as Sodium Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 12/01/03

Approved By:



Quality Control Supervisor