

Certificate of Analysis

Product Information:

Orion Number AC45CK
Description TURB CAL KIT AQ4500
Lot Number HU1 IR RATIO
Expiration Date 06/04
Method of Analysis Spectrophotometer

Actual Analysis:

<u>Analysis</u>	<u>Specification</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.89 TO 0.99	0.93	IR RATIO METHOD
2	9.6 TO 10.6	10.1	IR RATIO METHOD
3	104 TO 114	106	IR RATIO METHOD
4	693 TO 765	750	IR RATIO METHOD

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 01/08/04

Approved By:

Edward Henning

Quality Control Supervisor