

Certificate of Analysis

Product Information:

Orion Number	951211
Description	! AMMONIA PH ADJST ISA 475ML
Lot Number	HR1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	5.00 +/- 0.25 M Sodium Hydroxide / 0.05 +/- 0.005 Disodium EDTA

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	4.86	Moles/Liter Sodium Hydroxide
2	0.050	Moles/Liter Disodium EDTA

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Calcium Carbonate	915a	NA

Certificate Date: 09/11/03

Approved By:



Quality Control Supervisor