

Certificate of Analysis

Product Information:

Orion Number	911025
Description	PH 10 BUFFER IN 25 PACK.
Lot Number	HV2J
Expiration Date	05/05
Method of Analysis	Potentiometric
Acceptance Range	10.01 +/- 0.02 pH UNITS

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	10.011	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b

Certificate Date: 09/15/03

Approved By:



Quality Control Supervisor