

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	900002
<b>Description</b>	# INTERNAL FILL SOLN 5 X 50 ML
<b>Lot Number</b>	HQ1
<b>Expiration Date</b>	2 Years after Receipt of Product
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	0.70 M +/- 0.01 M Chloride

## Actual Analysis:


<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	0.71	Moles/Liter Potassium Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 10/06/03

**Approved By:**



**Quality Control Supervisor**