

Certificate of Analysis

Product Information:

Orion Number	080314
Description	ELECTROLYTE, D.O. PROBE 50ML BOTTLE
Lot Number	0017-3
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Conductivity

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	85 +/- 5	mS/cm at 25 deg. C

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Chloride	2202	NA

Certificate Date: 05/30/01

Approved By:



Quality Control Supervisor