

Certificate of Analysis

Product Information:

Orion Number	940906
Description	FLUORIDE STD, 475 ML
Lot Number	FV1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.1002	Moles/Liter Sodium Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 05/30/01

Approved By:

Edward Henning

Quality Control Supervisor