

Certificate of Analysis

Product Information:

Orion Number 700702
Description PURE WATER PH BUFFER A PH 6.97, 4 X 475 ML
Lot Number DX1
Expiration Date 03/31/01
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.0098	Moles/Liter Potassium chloride
2	6.974	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Potassium Hydrogen Phthalate	185	g

Certificate Date: 05/17/01

Approved By:

Edward Henning

Quality Control Supervisor