

Certificate of Analysis

Product Information:

Orion Number 910001
Description PH ELEC STORAGE SOLN, 475 ML
Lot Number FW1
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	3.02	Moles/Liter Potassium Chloride
2	6.71	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b
Sodium Chloride	2201	NA

Certificate Date: 04/30/01

Approved By:

Edward Henning

Quality Control Supervisor