

# Certificate of Analysis

## Product Information:

**Orion Number** 941706  
**Description** CHLORIDE ACTIVITY STD, 0.1M NaCl, 475ML  
**Lot Number** CP1  
**Expiration Date** 2 Years after Receipt of Product  
**Method of Analysis** Potentiometric

## Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u>      |
|-----------------|----------------|-----------------------------|
| 1               | 0.10017        | Moles/Liter Sodium Chloride |

## Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Sodium Chloride          | 2201         | NA            |

**Certificate Date:** 12/13/00

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

**Thermo**  
ELECTRON CORPORATION