

# Certificate of Analysis

## Product Information:

**Orion Number** 941706  
**Description** CHLORIDE ACTIVITY STD, 0.1M NaCl, 475ML  
**Lot Number** EU1  
**Expiration Date** 2 Years after Receipt of Product  
**Method of Analysis** Potentiometric

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.0997	Moles/Liter Sodium Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 06/15/00

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

**Thermo**  
ELECTRON CORPORATION