

Certificate of Analysis

Product Information:

Orion Number 841113
Description NA RECONDITIONING SOLN, 475 ML
Lot Number CY1
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.0992	Moles/Liter Ammonium Bifluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 10/26/99

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

Thermo
ELECTRON CORPORATION