

Certificate of Analysis

Product Information:

Orion Number 940909
Description TISAB II (ONE GALLON)
Lot Number DS2
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.065	ppm Fluoride
2	5.401	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA
Disodium Hydrogen Phosphate	186 II	f
Potassium Hydrogen Phthalate	185	g
Potassium Dihydrogen Phosphate	186 I	f

Certificate Date: 10/07/99

Approved By:

Edward Henning

Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

245967-001 Rev. B