

Certificate of Analysis

Product Information:

Orion Number 910001
Description PH ELEC STORAGE SOLN, 475 ML
Lot Number DU1
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	2.959	Moles/Liter Potassium Chloride
2	6.686	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Sodium Bicarbonate	191	a
Sodium Bicarbonate	191	b
Potassium Hydrogen Phthalate	185	g
Potassium Dihydrogen Phosphate	186 I	f
Disodium Hydrogen Phosphate	186 II	f

Certificate Date: 07/16/99

Approved By:



Quality Control Supervisor