

Certificate of Analysis

Product Information:

Orion Number 841111
Description NA ISA, 475 ML
Lot Number GT1
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|--------------------------------|
| 1 | 4.03 | Moles/Liter Ammonium Chloride |
| 2 | 4.02 | Moles/liter Ammonium Hydroxide |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Sodium Chloride | 2201 | NA |

Certificate Date: 07/18/02

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

245967-001 Rev. B

Thermo
ELECTRON CORPORATION