

Certificate of Analysis

Product Information:

Orion Number 700902
Description PURE WATER PH BUFFER C PH 9.15. 4 x 475 ML
Lot Number GP1
Expiration Date 11/30/03
Method of Analysis Potentiometric

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	9.183	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Dihydrogen Phosphate	186 I	f
Disodium Hydrogen Phosphate	186 II	f
Sodium Tetraborate Decahydrate	187	d

Certificate Date: 12/06/02

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

245967-001 Rev. B