

Certificate of Analysis

Product Information:

| | |
|---------------------------|-------------------------------------|
| Orion Number | 9107CB |
| Description | PH 7 BUFFER,CUBITAINER 19L |
| Lot Number | KU1 |
| Expiration Date | 6/08 |
| Method of Analysis | Potentiometric |
| Acceptance Range | 7.00 +/- 0.01 pH units at 25 deg. C |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 7.00 | pH @ 25 degrees C |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------------|--------------|---------------|
| Disodium Hydrogen Phosphate | 186 II | g |
| Potassium Dihydrogen Phosphate | 186 I | g |
| Sodium Bicarbonate | 191 | C |
| Sodium Carbonate | 191-I-c | NA |

Certificate Date: 11/22/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION