

Certificate of Analysis

Product Information:

| | |
|---------------------------|----------------------------------|
| Orion Number | 011008 |
| Description | COND STD 100 US 5X60ML |
| Lot Number | HV1 |
| Expiration Date | 05/31/05 |
| Method of Analysis | Conductivity |
| Acceptance Range | 100.0 uS/cm +/- 2.0 uS/cm @ 25.0 |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 100.5 | uS/cm at 25 deg. C |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Chloride | 2202 | NA |

Certificate Date: 06/04/03

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION