

Certificate of Analysis

Product Information:

Orion Number	934610
Description	NITRITE ISA, 475 ML
Lot Number	HS1
Expiration Date	8/05
Method of Analysis	Potentiometric
Acceptance Range	Ag+ 0.0100 M +/- 0.0005 M / pH 4.210 +/- 0.100

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
1	0.0098	Moles/Liter Silver Sulfate
2	4.179	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Potassium Hydrogen Phthalate	185	g
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b

Certificate Date: 08/26/03

Approved By:

Edward Henning

Quality Control Supervisor

Certificate of Analysis

Product Information:


Orion Number	934610
Description	NITRITE ISA, 475 ML
Lot Number	HS1
Expiration Date	8/05
Method of Analysis	Potentiometric
Acceptance Range	Ag+ 0.0100 M +/- 0.0005 M / pH 4.210 +/- 0.100

Actual Analysis:

<u>Analysis</u>	<u>Value</u>	<u>Unit of Measure</u>
-----------------	--------------	------------------------

Certificate Date: 08/26/03

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-8015 Don Fax
(978) 232-8031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION