

Certificate of Analysis

Product Information:

Orion Number	951007
Description	AMMONIA STANDARD 1000 P PM as N 475ML
Lot Number	IZ1
Expiration Date	1/06
Method of Analysis	Potentiometric
Acceptance Range	1000 ppm +/- 5 ppm Ammonium Chloride as Nitrogen

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	996.1	ppm as Nitrogen

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 01/13/04

Approved By:

Edward Henning

Quality Control Supervisor