

Certificate of Analysis

Product Information:

| | |
|---------------------------|---|
| Orion Number | 841113 |
| Description | NA RECONDITIONING SOLN, 475 ML |
| Lot Number | IZ1 |
| Expiration Date | 2 Years after Receipt of Product |
| Method of Analysis | Potentiometric |
| Acceptance Range | 0.100 M +/- 0.005 M Ammonium Bifluoride |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|---------------------------------|
| 1 | 0.096 | Moles/Liter Ammonium Bifluoride |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Fluoride | 2203 | NA |

Certificate Date: 01/14/04

Approved By:

Edward Henning

Quality Control Supervisor