

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	911025
<b>Description</b>	PH 10 BUFFER IN 25 PACK.
<b>Lot Number</b>	HS1E
<b>Expiration Date</b>	08/05
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	10.01 +/- 0.02 pH UNITS

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	10.013	pH

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	f
Potassium Dihydrogen Phosphate	186 I	f
Sodium Bicarbonate	191	b
Sodium Carbonate	192	b

**Certificate Date:** 01/26/04

**Approved By:**

*Edward Henning*

**Quality Control Supervisor**