

Certificate of Analysis

Product Information:

Orion Number	940909
Description	TISAB II (ONE GALLON)
Lot Number	IY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	pH 5.0 to 5.5/Fluoride < 0.10 ppm

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	5.155	pH
2	0.031	ppm Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA
Disodium Hydrogen Phosphate	186 II	g
Potassium Dihydrogen Phosphate	186 I	g
Potassium Hydrogen Phthalate	185	h
Sodium Bicarbonate	191	C
Sodium Carbonate	191	C

Certificate Date: 02/10/04

Approved By:

Edward Henning

Quality Control Supervisor

Certificate of Analysis

Product Information:


Orion Number	940909
Description	TISAB II (ONE GALLON)
Lot Number	IY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	pH 5.0 to 5.5/Fluoride < 0.10 ppm

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
-----------------	----------------	------------------------

Certificate Date: 02/10/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-8015 Don Fax
(978) 232-8031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION