

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	932011
<b>Description</b>	CALCIUM ISA, 475 ML
<b>Lot Number</b>	IY1
<b>Expiration Date</b>	2 Years after Receipt of Product
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	4.00 M +/- 0.20 M KCL

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.01	Moles/Liter Potassium Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 02/10/04

**Approved By:**



**Quality Control Supervisor**