

Certificate of Analysis

Product Information:

Orion Number	910001
Description	PH ELEC STORAGE SOLN, 475 ML
Lot Number	IY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	3.00 M +/- 0.10 M Potassium Chloride, 6.75 +/- 0.2 pH units]

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	2.98	Moles/Liter Potassium Chloride
2	6.70	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Disodium Hydrogen Phosphate	186 II	g
Potassium Dihydrogen Phosphate	186 I	g
Sodium Carbonate	191	C
Sodium Bicarbonate	191	C
Potassium Hydrogen Phthalate	185	h

Certificate Date: 02/26/04

Approved By:

Edward Henning

Quality Control Supervisor

Certificate of Analysis

Product Information:


Orion Number	910001
Description	PH ELEC STORAGE SOLN, 475 ML
Lot Number	IY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	3.00 M +/- 0.10 M Potassium Chloride, 6.75 +/- 0.2 pH units]

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
-----------------	----------------	------------------------

Certificate Date: 02/26/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-8015 Don Fax
(978) 232-8031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION