

Certificate of Analysis

Product Information:

| | |
|---------------------------|-------------------------------------|
| Orion Number | 040908 |
| Description | FLUORIDE STD, 10 PPM W/ TISAB |
| Lot Number | HO1 |
| Expiration Date | 12/05 |
| Method of Analysis | Potentiometric |
| Acceptance Range | 5.0 ppm +/- 0.08 ppm Fluoride (NaF) |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 5.04 | ppm Fluoride |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Fluoride | 2203 | NA |

Certificate Date: 12/23/03

Approved By:

Edward Henning

Quality Control Supervisor