

Certificate of Analysis

Product Information:

Orion Number	AC2009
Description	! AQUAFast II FLUORIDE KIT 50 TESTS 1ppm Fluoride Standard
Lot Number	HP1
Expiration Date	11/05
Method of Analysis	Potentiometric
Acceptance Range	1.00 ppm +/- 0.02 ppm Fluoride (NaF)

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.99	ppm Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 11/21/03

Approved By:

Edward Henning

Quality Control Supervisor