

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	040906
<b>Description</b>	# FLUORIDE STD. 1PPM W/ TISAB II
<b>Lot Number</b>	IY1
<b>Expiration Date</b>	2/06
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	0.50 +/- 0.01 ppm Fluoride

## Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.50	ppm Fluoride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

**Certificate Date:** 03/01/04

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

**Thermo**  
ELECTRON CORPORATION

244721-001 Rev. B