

Certificate of Analysis

Product Information:

Orion Number	941707
Description	CHLORIDE ACTIVITY STD, 100 PPM, 475ML
Lot Number	HP1
Expiration Date	11/05
Method of Analysis	Potentiometric
Acceptance Range	100.0 ppm +/- 0.5 ppm Sodium Chloride (NaCl) as Chloride (Cl-)

Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	99.81	ppm as Chloride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 12/03/03

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. C

Thermo
ELECTRON CORPORATION