

Certificate of Analysis

Product Information:

| | |
|---------------------------|---|
| Orion Number | 841113 |
| Description | NA RECONDITIONING SOLN, 475 ML |
| Lot Number | IW1 |
| Expiration Date | 2 Years after Receipt of Product |
| Method of Analysis | Potentiometric |
| Acceptance Range | 0.100 M +/- 0.005 M Ammonium Bifluoride |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|---------------------------------|
| 1 | 0.095 | Moles/Liter Ammonium Bifluoride |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Fluoride | 2203 | NA |

Certificate Date: 04/08/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION