

Certificate of Analysis

Product Information:

Orion Number	700902
Description	PURE WATER PH BUFFER C PH 9.15. 4 x 475 ML
Lot Number	IW1
Expiration Date	4/05
Method of Analysis	Potentiometric
Acceptance Range	pH 9.19 +/- 0.05 pH Units @ 25.0 C

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	9.20	pH @ 25 C

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Disodium Hydrogen Phosphate	186 II	g
Potassium Dihydrogen Phosphate	186 I	g
Sodium Tetraborate Decahydrate	187	d

Certificate Date: 04/20/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

244721-001 Rev. B