

Certificate of Analysis

Product Information:

Orion Number 951211
Description ! AMMONIA PH ADJST ISA 475ML
Lot Number IV1
Expiration Date 2 Years after Receipt of Product
Method of Analysis Potentiometric
Acceptance Range 5.00 +/- 0.25 M Sodium Hydroxide / 0.05 +/- 0.005 Disodium EDTA

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.86	Moles/Liter Sodium Hydroxide
2	0.050	Moles/Liter Disodium EDTA

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Calcium Carbonate	915a	NA

Certificate Date: 05/10/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

244721-001 Rev. B