

Certificate of Analysis

Product Information:

Orion Number	CODS10
Description	10000 PPM COD STANDARD
Lot Number	IV1
Expiration Date	5/06
Method of Analysis	Potentiometric
Acceptance Range	10000 ppm +/- 48 ppm Chemical Oxygen Demand

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	9976	ppm Chemical Oxygen Demand

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Hydrogen Phthalate	185	h

Certificate Date: 05/21/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

244721-001 Rev. B