

Certificate of Analysis

Product Information:

Orion Number	948207
Description	SULFATE STANDARD, 475 ML
Lot Number	IU1
Expiration Date	06-18-06
Method of Analysis	Potentiometric
Acceptance Range	0.1000 M +/-0.0005 M Sodium Sulfate

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	0.0999	Moles/Liter Sodium Sulfate

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Lead Nitrate	928	NA

Certificate Date: 07/01/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION