

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	941707
<b>Description</b>	CHLORIDE ACTIVITY STD, 100 PPM, 475ML
<b>Lot Number</b>	IU1
<b>Expiration Date</b>	6/06
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	100.0 ppm +/- 0.5 ppm Sodium Chloride (NaCl) as Chloride (Cl-)

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	99.84	ppm as Chloride

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 06/30/04

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

**Thermo**  
ELECTRON CORPORATION

244721-001 Rev. B