

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	951007
<b>Description</b>	AMMONIA STANDARD 1000 P PM as N 475ML
<b>Lot Number</b>	IT1
<b>Expiration Date</b>	07-06-06
<b>Method of Analysis</b>	Potentiometric
<b>Acceptance Range</b>	1000 ppm +/- 5 ppm Ammonium Chloride as Nitrogen

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	999.5	ppm as Nitrogen

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

**Certificate Date:** 07/09/04

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

**Thermo**  
ELECTRON CORPORATION

244721-001 Rev. B