

# Certificate of Analysis

## Product Information:

<b>Orion Number</b>	011005
<b>Description</b>	COND Std 111.9 mS/cm 5 pk-2oz.
<b>Lot Number</b>	IS1
<b>Expiration Date</b>	8/06
<b>Method of Analysis</b>	Conductivity
<b>Acceptance Range</b>	111.9 mS/cm +/- 0.5 mS/cm @ 25C +/- 0.2C

## Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	111.6	mS/cm at 25 deg. C

## Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Chloride	2202	NA

**Certificate Date:** 08/26/04

**Approved By:**



**Quality Control Supervisor**

Environmental Instruments  
Water Analysis

166 Cummings Center

Beverly, MA  
01915

(978) 232-6015 Dom Fax  
(978) 232-6031 Int'l Fax

www.thermo.com  
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

**Thermo**  
ELECTRON CORPORATION