

Certificate of Analysis

Product Information:

| | |
|---------------------------|----------------------------|
| Orion Number | 9107CB |
| Description | PH 7 BUFFER,CUBITAINER 19L |
| Lot Number | IQ1 |
| Expiration Date | 10/06 |
| Method of Analysis | Potentiometric |
| Acceptance Range | 7.00 +/- 0.01 pH units |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 7.003 | pH |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------------|--------------|---------------|
| Disodium Hydrogen Phosphate | 186 II | g |
| Potassium Dihydrogen Phosphate | 186 I | g |
| Sodium Bicarbonate | 191 | C |
| Sodium Carbonate | 191 | C |

Certificate Date: 10/27/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

Thermo
ELECTRON CORPORATION

244721-001 Rev. B