

Certificate of Analysis

Product Information:

Orion Number	CAC45FZ000
Description	4000 NTU FORMAZIN STANDARD, 500ML
Lot Number	IT1
Expiration Date	07-28-05
Method of Analysis	Nephelometric
Acceptance Range	4000 +/- 80 NTU

Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4020	NTU

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
NOT APPLICABLE	NA	NA

Certificate Date: 08/12/04

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis

166 Cummings Center

Beverly, MA
01915

(978) 232-6015 Dom Fax
(978) 232-6031 Int'l Fax

www.thermo.com
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. B

Thermo
ELECTRON CORPORATION