

Certificate of Analysis

Product Information:

Orion Number	951007
Description	AMMONIA STANDARD 1000 P PM as N 475ML
Lot Number	JU1
Expiration Date	06/07
Method of Analysis	Potentiometric
Acceptance Range	1000 ppm +/- 5 ppm Ammonium Chloride as Nitrogen

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	998.7	ppm as Nitrogen

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA

Certificate Date: 06/30/05

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. C

Thermo
ELECTRON CORPORATION