

Certificate of Analysis

Product Information:

Orion Number	910060
Description	PH STORAGE SOLUTION, 5X60 ML
Lot Number	KY1
Expiration Date	2/08
Method of Analysis	Potentiometric
Acceptance Range	3.00 M +/- 0.1 M KCL / 6.75 +/- 0.2 pH units

Actual Analysis:


<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	3.045	Moles/Liter Potassium Chloride
2	6.68	pH

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Sodium Chloride	2201	NA
Disodium Hydrogen Phosphate	186 II	g
Potassium Dihydrogen Phosphate	186 I	g
Sodium Bicarbonate	191	C
Sodium Carbonate	191	C

Certificate Date: 02/07/06

Approved By:



Quality Control Supervisor

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. C

Thermo
ELECTRON CORPORATION