

Certificate of Analysis

Product Information:

| | |
|---------------------------|--|
| Orion Number | 700902 |
| Description | PURE WATER PH BUFFER C PH 9.15. 4 x 475 ML |
| Lot Number | KZ1 |
| Expiration Date | 1/07 |
| Method of Analysis | Potentiometric |
| Acceptance Range | pH 9.19 +/- 0.05 pH Units @ 25.0 C |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|------------------------|
| 1 | 9.22 | pH @ 25 C |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------------|--------------|---------------|
| Disodium Hydrogen Phosphate | 186-II-g | NA |
| Potassium Dihydrogen Phosphate | 186-I-g | NA |
| Sodium Tetraborate Decahydrate | 187 | d |

Certificate Date: 02/15/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION