

Certificate of Analysis

Product Information:

Orion Number	951211
Description	! AMMONIA PH ADJST ISA 475ML
Lot Number	KY1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	5.00 +/- 0.25 M Sodium Hydroxide / 0.05 +/- 0.005 Disodium EDTA

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	4.95	Moles/Liter Sodium Hydroxide
2	0.052	Moles/Liter DiSodium EDTA

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Calcium Carbonate	915a	NA
Sodium Chloride	2201	NA

Certificate Date: 02/24/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION