

Certificate of Analysis

Product Information:

| | |
|---------------------------|---------------------------|
| Orion Number | 940906 |
| Description | FLUORIDE STD, 475 ML |
| Lot Number | KT1 |
| Expiration Date | 7/08 |
| Method of Analysis | Potentiometric |
| Acceptance Range | 0.1000 M +/- 0.0005 M NaF |

Actual Analysis:

| <u>Analysis</u> | <u>Results</u> | <u>Unit of Measure</u> |
|-----------------|----------------|-----------------------------|
| 1 | 0.0998 | Moles/Liter Sodium Fluoride |

Traceability:

| <u>N.I.S.T. Chemical</u> | <u>SRM #</u> | <u>Sublot</u> |
|--------------------------|--------------|---------------|
| Potassium Fluoride | 2203 | NA |

Certificate Date: 07/28/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION