

Certificate of Analysis

Product Information:

Orion Number	930711
Description	NITRATE ISA, 475 ML
Lot Number	KR1
Expiration Date	2 Years after Receipt of Product
Method of Analysis	Potentiometric
Acceptance Range	2.00 M +/- 0.10 M (NH ₄) ₂ SO ₄

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	2.01	Moles/Liter Ammonium Sulfate

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Lead Nitrate	928	NA

Certificate Date: 09/14/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION