

Certificate of Analysis

Product Information:

Orion Number	940907
Description	# FLUORIDE STD 475 ML 100 PPM
Lot Number	KR1
Expiration Date	9/08
Method of Analysis	Potentiometric
Acceptance Range	100.0 ppm +/- 0.5 ppm as F-

Actual Analysis:

<u>Analysis</u>	<u>Results</u>	<u>Unit of Measure</u>
1	100.1	ppm Fluoride

Traceability:

<u>N.I.S.T. Chemical</u>	<u>SRM #</u>	<u>Sublot</u>
Potassium Fluoride	2203	NA

Certificate Date: 10/06/06

Approved By:



Quality Representative

Environmental Instruments
Water Analysis Instruments

166 Cummings Center

Beverly, MA
01915

(978) 232-6000
(978) 232-6015 fax

www.thermo.com/water
(800) 225-1480 Tech Info

Analyze • Detect • Measure • Control™

244721-001 Rev. D

Thermo
ELECTRON CORPORATION